

CERTIFICATE OF ACCREDITATION

TANZANIA BUREAU OF STANDARDS (METROLOGY LABORATORY)

Established by the Standards Act No. 2 of 2009

Facility Accreditation Number: CAL-12 002

is a SADCAS accredited Calibration Laboratory
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,
Annexure "A", bearing the above accreditation number for

PRESSURE METROLOGY

The facility is accredited in accordance with the recognized International Standard

ISO/IEC 17025:2017

*The accreditation demonstrates technical competency for a defined scope and the operation
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States
as a multi-economy accreditation body*

Mrs Pinkie J Malebe
For SADCAS Chief Executive Officer

Effective Date (Issue No: 1): 13 December 2022
Certificate Expires: 15 February 2026

ANNEXURE A
SCHEDULE OF ACCREDITATION
PRESSURE METROLOGY

Laboratory Accreditation Number: CAL- (ISO/IEC 17025:2017)

<p><u>Permanent Address of Laboratory</u> Tanzania Bureau of Standards Metrology Laboratory Morogoro / Sam Nujoma Road, Ubungo Dar es Salaam Tanzania</p> <p><u>Postal Address</u> P O Box 9524 Dar es Salaam Tanzania</p> <p><u>Tel</u> : +255 22 2450206 <u>Cell</u> : +255 784 806 143 <u>Fax</u> : +255 22 245 0959 <u>Email</u> : joseph.mahilla@tbs.go.tz</p>	<p><u>Technical Signatories</u> : Mr J Yarrot Mr E Jinasa</p> <p><u>Nominated Representative</u> : Mr J J Mahilla</p> <p><u>Issue No</u> : 01 <u>Date of Issue</u> : 13 December 2022 <u>Expiry Date</u> : 15 February 2026</p>
---	--

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)
3.1	Pressure Gauges			At TBS and Onsite
3.1.1	Gas Medium • Pressure Gauges	Internal: MET-PV-02 (Calibration using pressure pump) Reference: SADCAS TR 22	0 bar to 20 bar	0.01 bar
3.1.2	Liquid Medium • Pressure Gauges	Internal: MET-PV-01 (Calibration using pressure balance) Reference: SADCAS TR 22	0 bar to 100 bar 100 bar to 200 bar 200 bar to 300 bar 300 bar to 400 bar 400 bar to 600 bar	At TBS 0.02 bar 0.03 bar 0.05 bar 0.05 bar 0.05 bar

Laboratory Accreditation No: CAL-12 002 (ISO/IEC 17025:2017)

Issue No: 01

Date of Issue: 13 December 2022

Date of Expiry: 15 February 2026

				At TBS and Onsite
3.1.3	Liquid Medium • Pressure Gauges(onsite)	Internal: MET-PV-03 (Calibration using pressure pump) Reference: SADCAS TR 22	0 bar to 100 bar 100 bar to 200 bar 200 bar to 300 bar 300 bar to 400 bar 400 bar to 600 bar	0.03bar 0.04 bar 0.08 bar 0.08 bar 0.08 bar

Original date of accreditation: 04 November 2010

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.

Pinkie J Malebe
SADCAS Technical Manager