

# CERTIFICATE OF ACCREDITATION

## CALTES COMPANY LIMITED - TANZANIA

*Company Registration No: 141005006*

**Facility Accreditation Number: CAL-5 002**

is a SADCAS accredited Calibration Laboratory  
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,  
Annexure "A", bearing the above accreditation number for

### **FORCE METROLOGY**

*The facility is accredited in accordance with the recognized International Standard*

### **ISO/IEC 17025:2017**

*The accreditation demonstrates technical competency for a defined scope and the operation  
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and  
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States  
as a multi-economy accreditation body*

---

**Eve Christine Gadzikwa**  
**SADCAS Chief Executive Officer**

**Effective Date (Issue No: 1): 31 August 2023**  
**Certificate Expires: 30 August 2028**

**ANNEXURE A**  
**SCHEDULE OF ACCREDITATION**  
**FORCE METROLOGY**

**Laboratory Accreditation Number: CAL-5 002 (ISO/IEC 17025:2017)**

<p><b><u>Permanent Address of Laboratory</u></b> Caltes Company Limited, Mbezi Malamba Mawili – Kinyerezi Road, Dar es Salaam Tanzania</p> <p><b><u>Postal Address</u></b> P O Box 10051 Dar es Salaam Tanzania</p> <p><b><u>Tel</u></b> : +255 754 408 014 <b><u>Cell</u></b> : +255 719 992 296 <b><u>Email</u></b> : <a href="mailto:b.shija@caltes.co.tz">b.shija@caltes.co.tz</a>; <a href="mailto:info@caltes.co.tz">info@caltes.co.tz</a>;</p>	<p><b><u>Technical Signatories</u></b> : Mr Barnabas M. Shija (All Items)</p> <p><b><u>Nominated Representative</u></b> : Mr Said H. Bussara</p> <p><b><u>Issue No</u></b> : 01 <b><u>Date of Issue</u></b> : 31 August 2023 <b><u>Expiry Date</u></b> : 30 August 2028</p>			
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)
1	Compression Machine (Uniaxial Testing Machine)	Internal: CTC-FORC:001	0 – 400 kN 400 – 1000 kN 1000 – 2000 kN	<i>At Caltes and On-site</i> 0,30 % 0,30% 0,24%

Original date of accreditation: 31 August 2023

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%.

---

**Pinkie J Malebe**  
**SADCAS Technical Manager**