

# CERTIFICATE OF ACCREDITATION

## **GOLDAIR TECHNOLOGIES PRIVATE LIMITED**

*Company Registration No. 8272/2021*

**Facility Accreditation Number: CAL-8 015**

is a SADCAS accredited Calibration Laboratory  
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,  
Annexure "A", bearing the above accreditation number for

## **MASS METROLOGY**

*The facility is accredited in accordance with the recognized International Standard*

## **ISO/IEC 17025:2017**

*The accreditation demonstrates technical competency for a defined scope and the operation  
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and  
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States  
as a multi-economy accreditation body*

---

**Eve Christine Gadzikwa**  
**SADCAS Chief Executive Officer**

**Effective Date (Issue No: 1): 29 August 2025**

**Certificate Expires: 28 August 2030**

## ANNEXURE A

### SCHEDULE OF ACCREDITATION

#### MASS METROLOGY

Laboratory Accreditation Number: **CAL-8 015 (ISO/IEC 17025:2017)**

<b>Permanent Address of Laboratory</b> Goldair Technologies 197 Longcheng Plaza Harare Zimbabwe  <b>Postal Address</b> 197 Longcheng Plaza Harare Zimbabwe  <b>Tel</b> : +263 24 271 1743 <b>Cell</b> : +263 77 210 7300 <b>Email</b> : <a href="mailto:bchibaya@gmail.com">bchibaya@gmail.com</a> <b>Email</b> : <a href="mailto:admin@goldairtechnologies.co.zw">admin@goldairtechnologies.co.zw</a>		<b>Technical Signatories</b> : Mr B Chibaya (All items)  <b>Nominated Representative</b> : Mr B Chibaya  <b>Issue No</b> : 01 <b>Date of Issue</b> : 29 August 2025 <b>Expiry Date</b> : 28 August 2030			
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	
				At Goldair	Onsite
1	Mass pieces	Internal: <i>GAP 027</i> Reference: <i>SADCAS TR 16;</i> <i>SADCAS TR 15;</i> <i>OIML R111-1</i>	1 mg to 100 mg 100 mg to 100 g 100 g to 500 g 500 g to 5000 g	0.07 mg 0.6 mg 3 mg 30 mg	0.08 mg 0.6 mg 3.5 mg 30 mg
2	Weighing Instruments • Digital self- indicating	Internal: <i>GAP 027</i> Reference: <i>SADCAS TR 16;</i> <i>SADCAS TR 15;</i> <i>OIML R76</i>	1 mg to 200 g 200 g to 500 g 500 g to 5000 g	0.2 mg 3 mg 30 mg	0.5 mg 5 mg 50 mg

Original date of accreditation: 29 August 2025

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%.

\_\_\_\_\_  
**Pinkie J Malebe**  
**SADCAS Technical Manager**