

CERTIFICATE OF ACCREDITATION

GOLDAIR TECHNOLOGIES PRIVATE LIMITED

Company Registration No. 8272/2021

Facility Accreditation Number: CAL-9 007

is a SADCAS accredited Calibration Laboratory
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,
Annexure "A", bearing the above accreditation number for

VOLUME METROLOGY

The facility is accredited in accordance with the recognized International Standard

ISO/IEC 17025:2017

*The accreditation demonstrates technical competency for a defined scope and the operation
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States
as a multi-economy accreditation body*

Eve Christine Gadzikwa
SADCAS Chief Executive Officer

Effective Date (Issue No: 1): 29 August 2025
Certificate Expires: 28 August 2030

ANNEXURE A
SCHEDULE OF ACCREDITATION
VOLUME METROLOGY

Laboratory Accreditation Number: CAL-9 007 (ISO/IEC 17025:2017)

<p>Permanent Address of Laboratory Goldair Technologies 197 Longcheng Plaza Harare Zimbabwe</p> <p>Postal Address 197 Longcheng Plaza Harare Zimbabwe</p> <p>Tel : +263 24 271 1743 Cell : +263 77 210 7300 Email : bchibaya@gmail.com; admin@goldairtechnologies.co.zw</p>		<p>Technical Signatories : Mr B Chibaya (All items)</p> <p>Nominated Representative : Mr B Chibaya</p> <p>Issue No : 01 Date of Issue : 29 August 2025 Expiry Date : 28 August 2030</p>			
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	
				At Goldair	Onsite
1	Piston Pipettes	Internal: GAP022 Reference: <i>Measurement Good Practice Guide No. 69, NPL, UK</i> <i>EURAMET Calibration Guide No. 19 Version 3.0 (09/2018);</i> <i>SADCAS TR 19</i>	5 µℓ to 100 µℓ 100 µℓ to 250 µℓ 250 µℓ to 500 µℓ 500 µℓ to 1 000 µℓ 1 000 µℓ to 3 000 µℓ 3 000 µℓ to 5 000 µℓ 5 000 µℓ to 10 000 µℓ	0.6 µℓ 0.8 µℓ 2.0 µℓ 3.0 µℓ 5 µℓ 12 µℓ 30 µℓ	0.8 µℓ 1.0 µℓ 2.5 µℓ 3.5 µℓ 8 µℓ 15 µℓ 35 µℓ
2	Glassware	Internal: GAP022 Reference: <i>SADCAS TR 19</i> <i>EURAMET Calibration Guide No. 19 Version 3.0 (09/2018);</i> <i>ISO 8106-2004 (E)</i>	1 mℓ to 5 mℓ 5 mℓ to 50 mℓ 50 mℓ to 250 mℓ 250 mℓ to 1000 mℓ 1 000 mℓ to 5 000 mℓ	0.05 mℓ 0.10 mℓ 0.20 mℓ 0.50 mℓ 2.0 mℓ	0.06 mℓ 0.12 mℓ 0.22 mℓ 0.53 mℓ 2.2 mℓ

Original date of accreditation: 29 August 2025

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.

Pinkie J Malebe
SADCAS Technical Manager