

# CERTIFICATE OF ACCREDITATION

## PANGOLIM PRESTAÇÃO DE SERVIÇOS LIMITADA

*Company Registration No. 2019-033*

**Facility Accreditation Number: CAL-14 013**

is a SADCAS accredited Calibration Laboratory  
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,  
Annexure "A", bearing the above accreditation number for

### TEMPERATURE METROLOGY

*The facility is accredited in accordance with the recognized International Standard*

### ISO/IEC 17025:2017

*The accreditation demonstrates technical competency for a defined scope and the operation  
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and  
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States  
as a multi-economy accreditation body*

---

**Eve Christine Gadzikwa**  
**SADCAS Chief Executive Officer**

**Effective Date (Issue No: 1): 29 September 2025**  
**Certificate Expires: 28 September 2030**

**ANNEXURE A**  
**SCHEDULE OF ACCREDITATION**  
**TEMPERATURE METROLOGY**

**Laboratory Accreditation Number: CAL-14 013 (ISO/IEC 17025:2017)**

<p><b><u>Permanent Address of Laboratory</u></b> Pangolim Prestação de Serviços Limitada Unit 6, Bizpark Viana Cacuaco-Benfica Luanda, Angola Angola</p> <p><b><u>Postal Address</u></b> Unit 6, Bizpark Viana Cacuaco-Benfica Luanda, Angola</p> <p><b><u>Tel</u></b> : +244 93 598 6940 <b><u>Cell</u></b> : +44 79 3583 4346 (UK) <b><u>Email</u></b> : <a href="mailto:greg.dinkelman@kairosmetering.com">greg.dinkelman@kairosmetering.com</a>;</p>	<p><b><u>Technical Signatories</u></b> : Mr I Aggrey (All items)</p> <p><b><u>Nominated Representative</u></b> : Mr G Dinkelman</p> <p><b><u>Issue No</u></b> : 01 <b><u>Date of Issue</u></b> : 29 September 2025 <b><u>Expiry Date</u></b> : 28 September 2030</p>
--	--

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	
				At Pangolim	Onsite
1	Temperature Sensors and Indicators (PT100@385)	Internal: IT.LAB.011 Reference: SADCAS TR26	-35 °C to 140 °C	0.05 °C	0.1 °C
2	Temperature Block/Field Dry well	IT.LAB.011B; SADCAS TR26 EURAMET cg-13	-35 °C to 140 °C	0.2 °C	0.2 °C

Original date of accreditation: 29 September 2025

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%.

\_\_\_\_\_  
**Pinkie J Malebe**  
**SADCAS Technical Manager**