

CERTIFICATE OF ACCREDITATION

MEDICINES CONTROL AUTHORITY OF ZIMBABWE

*Established by an Act of Parliament, the Medicines and Allied
Substances Control amendment Act No 1 of 1996 Chapter 15:03*

Facility Accreditation Number: TEST-11 0002

is a SADCAS accredited Testing Laboratory
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,
Annexure "A", bearing the above accreditation number for

PHYSICAL TESTING

The facility is accredited in accordance with the recognized International Standard

ISO/IEC 17025:2017

*The accreditation demonstrates technical competency for a defined scope and the operation
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States
as a multi-economy accreditation body*

Eve Christine Gadzikwa
SADCAS Chief Executive Officer

Date of Renewal of Accreditation: 07 February 2026
Effective Date (Issue No: 1): 12 February 2026
Certificate Expires: 11 February 2031

ANNEXURE A

SCHEDULE OF ACCREDITATION

PHYSICAL TESTING

Laboratory Accreditation Number: **TEST-11 0002 (ISO/IEC 17025:2017)**

<p>Permanent Address of Laboratory Medicines Control Authority of Zimbabwe 106 Baines Avenue Harare Zimbabwe</p> <p>Postal Address P O Box 10559 Harare Zimbabwe</p> <p>Tel : +263 4 736 981/5 Cell : +263 77 214 5191 Email : mcaz@mcaz.co.zw</p>		<p>Technical Signatories :</p> <p>Ms L Nhukarume (All Methods) Mr N O Shonhiwa (All Methods except Total Lubricants)</p> <p>Nominated Representative : Ms A Chikowore</p> <p>Issue No : 1 Date of Issue : 12 February 2026 Expiry Date : 11 February 2031</p>
MATERIALS/PRODUCTS TESTED	TYPES OF TESTS/PROPERTIES MEASURED, RANGE OF MEASUREMENT	STANDARD SPECIFICATIONS, EQUIPMENT/TECHNIQUES USED
Male Rubber Latex Condoms	<p>Dimension Testing</p> <ul style="list-style-type: none"> - Determination of Width - Determination of Length - Determination of Thickness <p>Freedom from Holes</p> <p>Testing for package integrity</p> <p>Determination of Bursting Volume and Pressure</p> <p>Determination of Total Lubricants</p>	<p>SOP MD-MT 028, MD-MT 006 / ISO 4074</p> <p>By Enersol automated Stainless Steel Ruler By Enersol automated Stainless Steel Mandrel By Enersol automated Thickness Gauge</p> <p>SOP MD-MT 030 / ISO 4074 by Enersol Ultra-fast Electrical Conductivity Tester</p> <p>SOP MD-MT 005/ISO 4074 Water Leak Test method by Enersol Visual Leaks Tester</p> <p>SOP MD-MT 031 / ISO 4074 by Enersol Package Seal Integrity Tester</p> <p>SOP MD-MT 003/ ISO 4074 By Enersol Air Inflation Testers</p> <p>SOP MD 012/ISO 4074</p>